

ANNEX II + III: TECHNICAL SPECIFICATIONS + TECHNICAL OFFER

Contract title: Supply of equipment for relevant institutions within the SoHo System in Serbia

p 1 /...

Publication reference: < NEAR/BEG/2021/EA-OP/0180>

Columns 1-2 should be completed by the contracting authority

Columns 3-4 should be completed by the tenderer

Column 5 is reserved for the evaluation committee

Annex III - the contractor's technical offer

The tenderers are requested to complete the template on the next pages:

- Column 2 is completed by the contracting authority shows the required specifications (not to be modified by the tenderer),
- Column 3 is to be filled in by the tenderer and must detail what is offered (for example the words 'compliant' or 'yes' are not sufficient)
- Column 4 allows the tenderer to make comments on its proposed supply and to make eventual references to the documentation

The eventual documentation supplied should clearly indicate (highlight, mark) the models offered and the options included, if any, so that the evaluators can see the exact configuration. Offers that do not permit to identify precisely the models and the specifications may be rejected by the evaluation committee.

The offer must be clear enough to allow the evaluators to make an easy comparison between the requested specifications and the offered specifications.

Unless otherwise specified, the requirements in these Technical Specifications are presented as a minimum standard which the offered goods must meet.

LOT 5 – Specialized Vehicles for Blood Transfusion Services

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
5.1	<p>VAN FOR TRANSPORT OF BLOOD TRANSFUSION TEAM WITH EQUIPMENT FOR SAFE TRANSPORT OF BLOOD</p> <p>Quantity: 11</p>			
	<ul style="list-style-type: none"> • Engine: diesel, 140 HP • Front or back wheel drive • Vehicle dimension: length 5400 mm, height 2500 mm, space volume 11 m2 • Vehicle layout: minimum 4 doors, 5+1 seats • Driver's and passenger airbags • ABS, ESP • Central locking • Front fog lights • Air conditioning for driving space and passenger compartment with cooling and heating options • Radio/Bluetooth/USB <p>Equipment:</p> <ul style="list-style-type: none"> • Mobile oxygen set • 2 blood donor chairs portable • 2 crew table-portable • 8 portable chairs for crew • Cupboard for medical supplies • Blood refrigerator portable, temperature range +40°C to -24°C, 80 l capacity, 	<p><i>Manufacturer's name:</i></p> <p><i>Product type, model:</i></p> <p><i>Specifications:</i></p>		

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	voltage 12V / 230V <ul style="list-style-type: none"> • Dispenser for hand disinfection • Medical waste bin 			

Installation and other services for Lot No5

Specifications Required		Specifications Offered	Notes, remarks, ref to documentation	Evaluation Committee's notes YES/NO
Manuals	Instruction manuals in Serbian and English			
Vehicle marking	Each vehicle must be marked on both sides with durable/UV resistant plastic stickers, the pattern must be agreed after the contract signature with the Contracting Authority.			
Certificates and documentation	Medical equipment, tools and consumables shall be in possession of European conformity marking (CE Mark), according to 93/42/EEC and 2007/47 directives, and applicable IEC/ISO 60601 and 80601 family standards. Manufacturer of the goods shall be ISO 13485 and/or ISO 9001 certified.			
Warranty	365 days after provisional acceptance in accordance with the conditions laid down in Article 32 of the General Conditions.			

Annex II + III: Technical Specifications + Technical Offer - part II – Place of delivery/Acceptance

Item	ARTICLE	Place of acceptance
-------------	----------------	----------------------------

Item 1	<p>VAN FOR TRANSPORT OF BLOOD TRANSFUSION TEAM WITH EQUIPMENT FOR SAFE TRANSPORT OF BLOOD</p>	<p><i>For the Transfusion Establishment, exact number the place of acceptances is 7:</i></p> <ol style="list-style-type: none"> <i>1. Blood Transfusion Institute of Serbia, Sveti Sava str.39, 11000 Belgrade;</i> <i>2. Blood Transfusion Institute Nis, Bulevar Dr Zorana Djindjica str.48, 18000 Nis;</i> <i>3. Institute for Blood Transfusion of Vojvodina, Hajduk Veljka str. 9a, 21000 Novi Sad;</i> <i>4. Institute for Blood Transfusion Kragujevac, Zmaj Jovina str. 30, 34000 Kragujevac;</i> <i>5. General Hospital Uzice, Miloša Obrenovića str.17, 31000 Užice;</i> <i>6. General Hospital Subotica, Izvorska str. 3, 24000 Subotica;</i> <i>7. KBC Zemun, Vukova str. 9, 11080 Zemun;</i>
--------	--	--